

温度測定記録紙

No.

測定年月日：2015 / 10 / 22

測定者：藤田 高士

観測地点：P07

温度計番号

標準：~~A022~~ ~~5~~ A022

アルコール：A021

温水 ~~冷水~~

経過時間 (分)	温水槽の温度 (°C)	
	標準温度計	アルコール温度計
0 (15:22)	26.1 26.1 26.1	25.9 25.9 25.9
1	26.1 26.1 26.1	26.0 26.0 26.0
2	26.1 26.1 26.1	26.0 26.0 26.0
3	26.1 26.1 26.1	26.0 26.0 26.0
4	26.1 26.1 26.1	26.0 26.0 26.0
5	26.1 26.1 26.1	26.0 26.0 26.0
6	26.1 26.1 26.1	26.0 26.0 26.0
7	26.1 26.1 26.1	26.0 26.0 26.0
8	26.1 26.1 26.1	26.0 26.0 26.0
9	26.1 26.1 26.1	26.0 26.0 26.0
10	26.1 26.1 26.1	26.0 26.0 26.0
11	26.1 26.1 26.1	26.0 26.0 26.0
12	26.1 26.1 26.1	26.0 26.0 26.0
13	26.1 26.1 26.1	26.0 26.0 26.0
14	26.1 26.1 26.1	25.9 25.9 25.9
15	26.1 26.1 26.1	25.9 25.9 25.9
16	26.1 26.1 26.1	25.9 25.9 25.9
17	26.1 26.1 26.1	25.9 25.9 25.9
18	26.1 26.1 26.1	25.9 25.9 25.9
19	26.1 26.1 26.1	26.0 26.0 26.0
20	26.1 26.1 26.1	26.0 26.0 26.0
21	26.1 26.1 26.1	26.0 26.0 26.0
22	26.1 26.2	26.0 26.0 26.0
23	26.1 26.1	26.0 26.0 26.0
24	26.1 26.2	26.0 26.0 26.0
25	26.1 26.2	26.0 26.0 26.0
26	26.1 26.2	26.0 26.0 26.0
27	26.1 26.1	26.0 26.0 26.0
28	26.1 26.1	26.0 26.0 26.0
29	26.1 26.2	26.0 26.0 26.0
30	26.1 26.2	26.0 26.0 26.0

冷水 ~~温水~~

経過時間 (分)	冷水槽の温度 (°C)	
	標準温度計	アルコール温度計
0	26.1 6.5	25.9 0.9
1	26.1 6.3	26.0 0.8
2	26.1 6.2	26.0 0.8
3	26.1 6.2	26.0 0.8
4	26.1 6.2	26.0 0.8
5	26.1 6.2	26.0 0.8
6	26.1 6.1	26.0 0.8
7	26.1 6.1	26.0 0.8
8	26.1 6.1	26.0 0.8
9	26.1 6.1	26.0 0.8
10	26.1 6.1	26.0 0.7
11	26.1 6.1	26.0 0.7
12	26.1 6.1	26.0 0.7
13	26.1 6.1	26.0 0.7
14	26.1 6.1	25.9 0.7
15	26.1 6.1	25.9 0.7
16	26.1 6.1	25.9 0.7
17	26.1 6.1	25.9 0.7
18	26.1 6.1	25.9 0.7
19	26.1 6.1	26.0 0.7
20	26.1 6.1	26.0 0.7
21	26.1 6.1	26.0 0.7
22	26.1 6.1	26.0 0.7
23	26.2 6.1	26.0 0.7
24	26.2 6.1	26.0 0.7
25	26.2 6.1	26.0 0.7
26	26.2 6.0	26.0 0.6
27	26.2 6.0	26.0 0.6
28	26.2 6.0	26.0 0.6
29	26.2 6.0	26.0 0.6
30	26.2 6.0	26.0 0.6