

温度測定記録紙

No. 1

測定年月日： 2015. 10. 22

測定者： 島 史佳

観測地点： p16

温度計番号

標準： A014

アルコール： A027

経過時間 (分)	温水槽の温度 (°C)	
	標準温度計	アルコール温度計
0	26.1	26.0
1	26.1	26.0
2	26.1	26.0
3	26.1	26.0
4	26.1	26.0
5	26.1	26.0
6	26.1	26.0
7	26.1	26.0
8	26.1	26.0
9	26.1	26.0
10	26.1	26.0
11	26.1	26.0
12	26.1	26.0
13	26.1	26.0
14	26.1	26.0
15	26.1	26.0
16	26.1	26.0
17	26.1	26.0
18	26.1	25.9
19	26.0	25.9
20	26.0	25.9
21	26.0	25.9
22	26.0	25.9
23	26.0	25.9
24	26.0	25.9
25	26.0	25.9
26	26.1	26.0
27	26.0	26.0
28	26.1	26.0
29	26.0	25.9
30	26.0	25.9

経過時間 (分)	冷水槽の温度 (°C)	
	標準温度計	アルコール温度計
0	1.0	0.9
1	1.0	0.9
2	1.0	0.9
3	1.0	0.9
4	1.0	0.9
5	1.0	0.8
6	0.9	0.8
7	0.9	0.8
8	0.9	0.8
9	0.9	0.8
10	0.9	0.7
11	0.9	0.7
12	0.9	0.7
13	0.9	0.7
14	0.9	0.7
15	0.9	0.7
16	0.9	0.7
17	0.9	0.7
18	0.9	0.7
19	0.9	0.7
20	0.9	0.7
21	0.9	0.7
22	0.8	0.7
23	0.8	0.6
24	0.8	0.6
25	0.8	0.6
26	0.8	0.6
27	0.8	0.6
28	0.8	0.6
29	0.8	0.6
30	0.8	0.6